BIOZOL
Diagnostica Vertrieb GmbH, Oehleckerring 11-13
22419 Hamburg, Germany
Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99
E-Mail: info@biozol.de

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human UDP-glucuronic acid decarboxylase 1 (UXS1) ELISA Kit ASC-KTE60076

Artikelname	Human UDP-glucuronic acid decarboxylase 1 (UXS1) ELISA Kit
Artikelnummer	ASC-KTE60076
Hersteller Artikelnummer	KTE60076
Alternativnummer	ASC-KTE60076-48, ASC-KTE60076-96
Hersteller	Abbkine Scientific
Kategorie	Kits/Assays
Applikation	ELISA
Spezies Reaktivität	Human
Produktbeschreibung	This Human UDP-glucuronic acid decarboxylase 1 (UXS1) ELISA Kit employs a two-site sandwich ELISA to quantitate UXS1....
Detektionsbereich	Please inquire
Sensitivitaet	Please inquire
Tag	UXS1
NCBI	80146
UniProt	Q8NBZ7
Proben	Cell culture supernatantsSerumPlasmaOther biological fluids

Anwendungsbeschreibung	This Human UDP-glucuronic acid decarboxylase 1 (UXS1) ELISA Kit employs a two-site sandwich ELISA to quantitate UXS1 in samples. An antibody specific for UXS1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyUXS1 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for UXS1 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of UXS1 bound in the initial step. The color development is stopped and the intensity of the color is measured.