

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Mouse Ultrasensitivity Thyroxine (U-T4) ELISA Kit BYT-ORB1946770

Artikelname	Mouse Ultrasensitivity Thyroxine (U-T4) ELISA Kit
Artikelnummer	BYT-ORB1946770
Hersteller Artikelnummer	orb1946770
Alternativnummer	BYT-ORB1946770-48,BYT-ORB1946770-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Mouse
Produktbeschreibung	Mouse Ultrasensitivity Thyroxine (U-T4) ELISA Kit...
Detektionsbereich	0.79-50ng/mL
Sensitivitaet	0.47 ng/mL
Proben	serum, plasma, Tissue homogenate and Other biological samples
Target-Kategorie	U-T4

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Competitive-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with Mouse U-T4. During the reaction, Mouse U-T4 in the sample or standard competes with a fixed amount of Mouse U-T4 on the solid phase supporter for sites on the Biotinylated Detection Ab specific to Mouse U-T4. Excess conjugate and unbound sample or standard are washed away, and Avidin-Horseradish Peroxidase (HRP) conjugate are added to each micro plate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns from blue to yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm  $\pm$  2 nm. The concentration of Mouse U-T4 in tested samples can be calculated by comparing the OD of the samples to the standard curve