

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Rat Pituitary Adenylate Cyclase Activating Polypeptide 38 (PACAP-38) ELISA Kit BYT-ORB1946798

Artikelname	Rat Pituitary Adenylate Cyclase Activating Polypeptide 38 (PACAP-38) ELISA Kit
Artikelnummer	BYT-ORB1946798
Hersteller Artikelnummer	orb1946798
Alternativnummer	BYT-ORB1946798-48,BYT-ORB1946798-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Rat
Produktbeschreibung	Rat Pituitary Adenylate Cyclase Activating Polypeptide 38 (PACAP-38) ELISA Kit...
Detektionsbereich	7.82-500pg/mL
Sensitivitaet	4.69 pg/mL
Proben	serum, plasma, Tissue homogenate and Other biological samples
Target-Kategorie	PACAP-38

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Competitive-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with Rat PACAP-38. During the reaction, Rat PACAP-38 in the sample or standard competes with a fixed amount of Rat PACAP-38 on the solid phase supporter for sites on the Biotinylated Detection Ab specific to Rat PACAP-38. Excess conjugate and unbound sample or standard are washed away, and Avidin-Horseradish Peroxidase (HRP) conjugate are added to each micro plate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns from blue to yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The concentration of Rat PACAP-38 in tested samples can be calculated by comparing the OD of the samples to the standard curve