

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Melatonin Sulfate (MS) ELISA Kit BYT-ORB1946821

| | |
|--------------------------|---|
| Artikelname | Human Melatonin Sulfate (MS) ELISA Kit |
| Artikelnummer | BYT-ORB1946821 |
| Hersteller Artikelnummer | orb1946821 |
| Alternativnummer | BYT-ORB1946821-48,BYT-ORB1946821-96 |
| Hersteller | Biorbyt |
| Kategorie | Kits/Assays |
| Spezies Reaktivität | Human |
| Produktbeschreibung | Human Melatonin Sulfate (MS) ELISA Kit... |
| Detektionsbereich | 4.69-300ng/mL |
| Sensitivitaet | 2.81 ng/mL |
| Proben | serum, plasma, Tissue homogenate and Other biological samples |
| Target-Kategorie | MS |

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Competitive-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with Human MS. During the reaction, Human MS in the sample or standard competes with a fixed amount of Human MS on the solid phase supporter for sites on the Biotinylated Detection Ab specific to Human MS. Excess conjugate and unbound sample or standard are washed away, and Avidin-Horseradish Peroxidase (HRP) conjugate are added to each micro plate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns from blue to yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The concentration of Human MS in tested samples can be calculated by comparing the OD of the samples to the standard curve