

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Dehydroepiandrosterone Sulfate (DHEA-S) ELISA Kit BYT-ORB1946838

Artikelname	Human Dehydroepiandrosterone Sulfate (DHEA-S) ELISA Kit
Artikelnummer	BYT-ORB1946838
Hersteller Artikelnummer	orb1946838
Alternativnummer	BYT-ORB1946838-48,BYT-ORB1946838-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Human Dehydroepiandrosterone Sulfate (DHEA-S) ELISA Kit...
Detektionsbereich	15.63-1000ng/mL
Sensitivitaet	9.38 ng/mL
Proben	serum, plasma, Tissue homogenate and Other biological samples
Target-Kategorie	DHEA-S

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Competitive-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with Human DHEA-S. During the reaction, Human DHEA-S in the sample or standard competes with a fixed amount of Human DHEA-S on the solid phase supporter for sites on the Biotinylated Detection Ab specific to Human DHEA-S. Excess conjugate and unbound sample or standard are washed away, and Avidin-Horseradish Peroxidase (HRP) conjugate are added to each micro plate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns from blue to yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The concentration of Human DHEA-S in tested samples can be calculated by comparing the OD of the samples to the standard curve