

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Human Cross Linked C-telopeptide of Type III Collagen (CTX-III) ELISA Kit BYT-ORB1946852

Artikelname	Human Cross Linked C-telopeptide of Type III Collagen (CTX-III) ELISA Kit
Artikelnummer	BYT-ORB1946852
Hersteller Artikelnummer	orb1946852
Alternativnummer	BYT-ORB1946852-48,BYT-ORB1946852-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Human Cross Linked C-telopeptide of Type III Collagen (CTX-III) ELISA Kit...
Detektionsbereich	0.16-10ng/mL
Sensitivitaet	0.1 ng/mL
Proben	serum, plasma, Tissue homogenate and Other biological samples
Target-Kategorie	CTX-III

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Competitive-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with Human CTX-III. During the reaction, Human CTX-III in the sample or standard competes with a fixed amount of Human CTX-III on the solid phase supporter for sites on the Biotinylated Detection Ab specific to Human CTX-III. Excess conjugate and unbound sample or standard are washed away, and Avidin-Horseradish Peroxidase (HRP) conjugate are added to each micro plate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns from blue to yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm  $\pm$  2 nm. The concentration of Human CTX-III in tested samples can be calculated by comparing the OD of the samples to the standard curve