

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Dangerous Good - 4-component Contract Laboratory Program (CLP) ICP Interference Check Standard 1 500 mL Contains: HPS-CLP-INF-1-500

Article Name	Dangerous Good - 4-component Contract Laboratory Program (CLP) ICP Interference Check Standard 1 500 mL Contains:
Biozol Catalog Number	HPS-CLP-INF-1-500
Supplier Catalog Number	CLP-INF-1-500
Alternative Catalog Number	HPS-CLP-INF-1-500-500
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Interference Check Std 1 in 5% HNO ₃ , (µg/mL) Iron at 2,000, Aluminium, Calcium, Magnesium at 5,000 in 5% HNO ₃ . 12 months expiry date. Traceable to NIST SRM 31XX series. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited....