

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

7-component Initial Verification Graphite Furnace Calibration Standard 100 mL Contains: HPS-ICV-GFAA-100

Article Name	7-component Initial Verification Graphite Furnace Calibration Standard 100 mL Contains:
Biozol Catalog Number	HPS-ICV-GFAA-100
Supplier Catalog Number	ICV-GFAA-100
Alternative Catalog Number	HPS-ICV-GFAA-100-100
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Graphite Furn Cal Std in 2% HNO ₃ , (µg/mL) Arsenic, Lead, Selenium, Thallium at 2, Cadmium, Chromium, Silver at 0.4 in 2% HNO ₃ . 12 months expiry date. Traceable to NIST 31XX series. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accre...