

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Selenium (Se) ICP Standard at 1,000 µg/mL in 2% HNO₃ 500 mL HPS-100049-1-500

Article Name	Selenium (Se) ICP Standard at 1,000 µg/mL in 2% HNO ₃ 500 mL
Biozol Catalog Number	HPS-100049-1-500
Supplier Catalog Number	100049-1-500
Alternative Catalog Number	HPS-100049-1-500-500
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Se 1000 µg/mL in 2% HNO ₃ 18 months expiry date. Traceable to NIST SRM 3149. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited....