

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Vanadium (V) ICP Standard at 100 µg/mL in 2% HNO<sub>3</sub> 500 mL HPS-100-65-1-500

|                            |                                                                                                                                                                     |
|----------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Article Name               | Vanadium (V) ICP Standard at 100 µg/mL in 2% HNO <sub>3</sub> 500 mL                                                                                                |
| Biozol Catalog Number      | HPS-100-65-1-500                                                                                                                                                    |
| Supplier Catalog Number    | 100-65-1-500                                                                                                                                                        |
| Alternative Catalog Number | HPS-100-65-1-500-500                                                                                                                                                |
| Manufacturer               | High Purity Standards                                                                                                                                               |
| Category                   | Biochemikalien                                                                                                                                                      |
| Product Description        | V 100 µg/mL in 2% HNO <sub>3</sub> 12 months expiry date. Traceable to NIST SRM 3165. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited.... |