

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

AMPIPROBE Ureaplasma spp. / M. genitalium / M. hominis Assay Kit ENZ-ENZ-GEN209-0100

Article Name	AMPIPROBE Ureaplasma spp. / M. genitalium / M. hominis Assay Kit
Biozol Catalog Number	ENZ-ENZ-GEN209-0100
Supplier Catalog Number	ENZ-GEN209-0100
Alternative Catalog Number	ENZ-ENZ-GEN209-0100-100
Manufacturer	Enzo Life Sciences
Category	Kits/Assays
Species Reactivity	Bacteria
Product Description	AMPIPROBE Ureaplasma spp. / M. genitalium / M. hominis Assay Kit...
Sensitivity	In a validation study approved by the New York State Department of Health, the AMPIPROBE Ureaplasma spp. / M. genitalium / M. hominis Assay was determined to have the following percentage of Sensitivity: Ureaplasma spp. - 97.80% M. genitalium - 87.30 M.
Target	Ureaplasma spp. Mycoplasma genitalium Mycoplasma hominis
Application Notes	The AMPIPROBE Ureaplasma spp. / M. genitalium / M. hominis Assay is compatible with any properly calibrated qPCR thermal cycler capable to detect fluorescence decay. It has been validated for use on the QIAGEN Rotor-Gene Q