

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Human X-Ray Repair Cross Complementing 6 (XRCC6) CLIA Kit BIM-EKU09999

Article Name	Human X-Ray Repair Cross Complementing 6 (XRCC6) CLIA Kit
Biozol Catalog Number	BIM-EKU09999
Supplier Catalog Number	EKU09999
Alternative Catalog Number	BIM-EKU09999-96T
Manufacturer	Biomatik Corporation
Category	Kits/Assays
Species Reactivity	Human
Product Description	Human X-Ray Repair Cross Complementing 6 (XRCC6) CLIA Kit for detecting XRCC6 in tissue homogenates, other biological fluids. Species Reactivity: Human (Homo sapiens). Detection Range: 27.4-20,000pg/ml. Sensitivity: 11.9pg/ml. Format: 96-Well Plate. ...
Range	27.4-20,000pg/ml
Sensitivity	11.9pg/ml
NCBI	<a href="#">2547</a>
Form	96-Well Plate
Samples	Tissue Homogenates, Other Biological Fluids
Target	X-Ray Repair Cross Complementing 6 (XRCC6)