

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

CircuLex Human Chitotriosidase ELISA Kit MBL-CY-8074

Article Name	CircuLex Human Chitotriosidase ELISA Kit
Biozol Catalog Number	MBL-CY-8074
Supplier Catalog Number	CY-8074
Alternative Catalog Number	MBL-CY-8074
Manufacturer	MBL
Category	Kits/Assays
Application	ELISA
Species Reactivity	Hu
Product Description	CircuLex Human Chitotriosidase ELISA Kit...
Molecular Weight	Dilution factors need to be taken into consideration in calculating the human chitotriosidase concentration. Results exceeding human chitotriosidase level of 3.6 ng/ml should be repeated with diluted samples.
Range	Dilution factors need to be taken into consideration in calculating the human chitotriosidase concentration. Results exceeding human chitotriosidase level of 3.6 ng/ml should be repeated with diluted samples.
Sensitivity	better than 48.3 pg/ml of sample.
Target	CHITO