

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human Osteogenic Growth Peptide (OGP) ELISA Kit BYT-ORB1946848

| | |
|----------------------------|---|
| Article Name | Human Osteogenic Growth Peptide (OGP) ELISA Kit |
| Biozol Catalog Number | BYT-ORB1946848 |
| Supplier Catalog Number | orb1946848 |
| Alternative Catalog Number | BYT-ORB1946848-48,BYT-ORB1946848-96 |
| Manufacturer | Biorbyt |
| Category | Kits/Assays |
| Species Reactivity | Human |
| Product Description | Human Osteogenic Growth Peptide (OGP) ELISA Kit... |
| Range | 0.16-10ng/mL |
| Sensitivity | 0.1 ng/mL |
| Samples | serum, plasma, Tissue homogenate and Other biological samples |
| Target | OGP |

Application Notes

Application Notes: This ELISA kit uses the Competitive-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with Human OGP. During the reaction, Human OGP in the sample or standard competes with a fixed amount of Human OGP on the solid phase supporter for sites on the Biotinylated Detection Ab specific to Human OGP. Excess conjugate and unbound sample or standard are washed away, and Avidin-Horseradish Peroxidase (HRP) conjugate are added to each micro plate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns from blue to yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The concentration of Human OGP in tested samples can be calculated by comparing the OD of the samples to the standard curve